

An attempt of finding transfer function between standard 12 lead ECG system and EASI ECG system. Various regression techniques, like Least Angle Regression ("LARS"), LASSO, Forward Stagewise, linear regression, were used to find a set of transfer coefficients. Different tools were used (Matlab, R-Project, WEKA) for better problem investigation. Few transfer functions that successfully approximate EASI ECG system were obtained.