

The contribution deals with calibration of the 3D ultrasonic tomography system. The increased precision of signals TOF measuring is demonstrated to be obtained using oversampling in the frequency domain after choosing a properly designed chirp excitation pulse.